



NATIONAL TYPE EVALUATION PROGRAM

Certificate of Conformance

for Weighing and Measuring Devices

For:

Non-Computing Scale
Digital Electronic
Model: SJ Series
 n_{max} : 2 600
 e_{min} : 0.001 lb / 0.0005 kg / 0.02 oz
Capacity: 2.2 to 66 lb / 1 to 30 kg / 35 to 1058 oz
Platform Size: 7.5 in x 9.0 in
Accuracy Class: III

Submitted By:

A&D Engineering
1756 Automation Parkway
San Jose, CA 95131
Tel: 408-518-5112
Fax: 408-635-2312
Contact: Adnan Alam
Email: alam@andonline.com
Web site: www.andweighing.com

Standard Features and Options

Standard Features:

- Semi-automatic (push-button) Zero (SAZSM)
- Automatic Zero Tracking (AZT)
- Unit Conversion (kg, g, lb, oz)
- AC or DC Power
- Battery Saving Feature (Auto-shut off)
- Initial Zero Setting Mechanism (IZSM)
- Semi-automatic (push-button) Tare
- LCD Display
- Battery Power Optional

Load Cells Used:


- A&D Load-Cell Model (CC:10-001): 1ETTS103B (non-NTEP)
- A&D Load-Cell Model (CC: 10-001A1): LC184 (non-NTEP)
- HS Models = Stainless Steel Pan
- H Models = Plastic Pan (as part of the unit itself)


Specific Models, capacities and n_{max} :

Model	Capacity x d (lb)	Capacity x d (kg)	Capacity x d (g)	Capacity x d (oz)	n_{max}
SJ-1000H/HS	2.2 x 0.001	1 x 0.0005	1000 x 0.5	35 x 0.02	2200
SJ-2000H/HS	4.4 x 0.002	2 x 0.001	2000 x 1	70 x 0.05	2200
SJ-5000H/HS	11 x 0.005	5 x 0.002	5000 x 2	176 x 0.1	2500
SJ-12KH/HS	26 x 0.01	12 x 0.005	12000 x 5	423 x 0.2	2600
SJ-20KH/HS	44 x 0.02	20 x 0.01	20000 x 10	705 x 0.5	2200
SJ-30KH/HS	66 x 0.05	30 x 0.02	30000 x 20	1058 x 1	1500

Temperature Range: -10 °C to 40 °C (14 °F to 104 °F)

This device was evaluated under the National Type Evaluation Program and was found to comply with the applicable technical requirements of "NIST Handbook 44: Specifications, Tolerances and Other Technical Requirements for Weighing and Measuring Devices." Evaluation results and device characteristics necessary for inspection and use in commerce are on the following pages.


Stephen Benjamin
Chairman, NCWM, Inc.


Kurt Floren
Chairman, National Type Evaluation Program Committee
Issued: April 24, 2013

1135 M Street, Suite 110 / Lincoln, Nebraska 68508

The National Conference on Weights and Measures (NCWM) does not approve, recommend or endorse any proprietary product or material, either as a single item or as a class or group. Results shall not be used in advertising or sales promotion to indicate explicit or implicit endorsement of the product or material by the NCWM.



A&D Engineering

Non-Computing Scale / SJ Series

Application: Non computing scale used for general purpose weighing applications.

Identification: The required information is on an adhesive badge affixed on the side of the indicating element.

Sealing: The device is sealed on underside by means of a wire security seal threaded through two eyelets one on the calibration cover plate the other on scale casing, to prevent access to the calibration switch.

Test Conditions: This Certificate supersedes Certificate of Conformance Number 10-001 and is issued to include new Model number suffix HS which represents a new load cell and has a stainless steel pan. The emphasis of the evaluation was on the device design, operation, marking requirements and compliance with influence factor requirements. For the purpose of this evaluation, a model SJ-1000HS (2.2 lb x 0.001 lb), SJ-12KHS (26 lb x 0.01 lb), and SJ-30KHS (30 kg x 0.02 kg) were submitted. Several increasing/decreasing load and shift tests were conducted on each scale. The scales were tested over a temperature range of -10 °C to 40 °C (14 °F to 104 °F). A load of approximately one-half scale capacity was applied to each scale over 100 000 times. The scales were tested periodically during this time. Tests were also conducted with a power supply of 102 VAC to 132VAC, and 5.8 VDC to 10.0 VDC. Results of the previous test conditions are repeated below for reference.

Certificate of Conformance Number 10-001: The emphasis of the evaluation was on the device design, operation, marking requirements and compliance with influence factor requirements. For the purpose of this evaluation, a model SJ-1000H (2.2 lb x 0.001 lb), SJ-12KH (26 lb x 0.01 lb), and SJ-30KH (30 kg x 0.02 kg) were submitted. Several increasing/decreasing load and shift tests were conducted on each scale. The scales were tested over a temperature range of -10 °C to 40 °C (14 °F to 104 °F). A load of approximately one-half scale capacity was applied to each scale over 100 000 times. The scales were tested periodically during this time. Tests were also conducted with a power supply of 100 VAC to 130VAC, and 5.8 VDC to 10.0 VDC.

Evaluated By: A. McCoy (OH), J. Morrison (OH), M. Kelley (OH), T. Buck (OH), E. Matthews (OH) 10-001; C. Harris (OH) 10-001A1

Type Evaluation Criteria Used: *NIST Handbook 44 Specifications, Tolerances, and Other Technical Requirements for Weighing and Measuring Devices*, 2013 Edition. *NCWM Publication 14 Weighing Devices*, 2013 Edition.

Conclusion: The results of the evaluation and information provided by the manufacturer indicate the device complies with applicable requirements.

Information Reviewed By: J. Truex (NCWM) 10-001, 10-001A1

Examples of Device:

