



## NATIONAL TYPE EVALUATION PROGRAM

# Certificate of Conformance

for Weighing and Measuring Devices

**For:**

Non-Computing Scale  
 Digital Electronic  
 Model: WPB Series  
 $n_{max}$ : 3000  
 $e_{min}$ : (see table below)  
 Capacity: (see table below)  
 Platform: 220 cm x 180 cm (8.5 in x 7 in)  
 Accuracy Class: III

**Submitted By:**

Cardinal Scale Manufacturing Company, Inc.  
 102 East Daugherty Street  
 Webb City, MO 64870  
 Tel: 417-673-4631  
 Fax: 417-673-2153  
 Contact: Eric Golden  
 Email: [egolden@cardet.com](mailto:egolden@cardet.com)  
 Web site: [www.cardet.com](http://www.cardet.com)

**Standard Features and Options**

- Semi-Automatic Zero Setting Mechanism (Push Button)
- Initial Zero Setting Mechanism (IZSM)
- Automatic Zero Tracking (AZT)
- Semi-Automatic Tare (Push Button)
- Alphanumeric Display
- Customer Display (Dual)
- AC Power
- DC Power
- Sleep Mode
- Auto Shut Off
- Single Range
- Multi Range
- RS-232/USB

Model	Capacity (kg)	$e_{min}$ (g)	Capacity (lb)	$e_{min}$ (oz)
WPB-6	1.5 kg / 3 kg	0.5 g / 1 g	3 lb / 6 lb	0.02 oz / 0.05 oz
WPB-12	3 kg / 6 kg	1 g / 2 g	6 lb / 12 lb	0.05 oz / 0.1 oz
WPB-30	6 kg / 15 kg	2 g / 5 g	12 lb / 30 lb	0.1 oz / 0.2 oz
WPB-65	15 kg / 30 kg	5 g / 10 g	30 lb / 65 lb	0.2 oz / 0.5 oz

**Load Cell Used:** Zemic L6D (NTEP CC 11-012), or other NTEP certified metrological equivalent

Temperature Range: -10 °C to 40 °C (14 °F to 104 °F)

This device was evaluated under the National Type Evaluation Program and was found to comply with the applicable technical requirements of "NIST Handbook 44: Specifications, Tolerances and Other Technical Requirements for Weighing and Measuring Devices." Evaluation results and device characteristics necessary for inspection and use in commerce are on the following pages.

Hal Prince  
 Chairman, NCWM, Inc.

Craig VanBuren  
 Chair, NTEP Committee  
 Issued: July 24, 2020

1135 M Street, Suite 110 / Lincoln, Nebraska 68508

The National Conference on Weights and Measures (NCWM) does not approve, recommend or endorse any proprietary product or material, either as a single item or as a class or group. Results shall not be used in advertising or sales promotion to indicate explicit or implicit endorsement of the product or material by the NCWM.



**Cardinal Scale Manufacturing Company, Inc.**  
Non-Computing Scale / WPB Series

**Application:** For use in general purpose weighing applications.

**Identification:** A self-destructive sticker with the identification information is located on the side of the unit.

**Sealing:** Wire security seal threaded through two screws that cover a plate that grants access to the calibration switch.

**Test Conditions:** This Certificate is issued based on the following tests and information provided by the manufacturer. The emphasis of this evaluation was on the design, marking requirements, operation and compliance with influence factor requirements. Two multi-range scales were provided for this evaluation, a 3kg capacity and a 30kg capacity. These devices were tested over a temperature range of -10 °C to 40 °C (14 °F to 104 °F). They also had a load of approximately one-half capacity applied over 100,000 times with several increasing / decreasing load tests performed during this time. Several shift and load discrimination tests were conducted during the evaluation. The units under evaluation were tested with supply voltages from 102 to 132 VAC, 4.2 to 6.6 VDC and were also subjected to RFI testing.

**Evaluated By:** E. Morabito (NY)

**Type Evaluation Criteria Used:** *NIST Handbook 44 Specifications, Tolerances, and Other Technical Requirements for Weighing and Measuring Devices*, 2016 Edition. *NCWM Publication 14 Measuring Devices*, 2016 Edition.

**Conclusion:** The results of the evaluation and information provided by the manufacturer indicate the device complies with applicable requirements.

**Information Reviewed By:** D. Flocken (NCWM)

**Example(s) of Device:**



**Calibration access via lead wire seal**